Certificate of Calibration

Certificate Procedure Number: EDCQP200-4.11.5

Environmental Devices Corporation certifies the Haz-Dust Particulate Monitors are calibrated gravimetrically against the specifications and protocols set forth in NIOSH method 0600 and/or 0500. Calibration is NIST traceable and conforms to original published specifications of +/- 10%.

Calibration Dust Specifications are determined with a NIST traceable Coulter Mutisizer, ISO12103 –1 A2 Fine Test Dust and primary Flow Standard: LFE774300.

Our quality system standard meets the requirements of ANSI/NCSLZ540.1 and ASQC standard ISO 9001, MIL-STD 45662A, and customer's specifications if requested.

Calibration Test Dust Particulate	**Cumulative Volume Numeric Data**
Micron Size	% Less Than
1	2.9
2	11
3	19.6
4	27.7
5	34.6
7	43.6
10	52.1
20	70.7
40	89.2
80	99.8
120	100

Temperature = 22°C
Relative Humidity = 30%
Atmospheric Pressure = 760 mmHg
Measurement Uncertainty Estimated @ 95% Confidence Level (k=2)

Model	Serial Number	Calibration Date	Next Calibration Due Date
EPAM-5000	0714497	July 1, 2014	July 2015

Calibration Span Accessory if purchased	K= 18.4 mg/m3	Model: CS-105

Technician	**Supervisor**
Matt Gosslin	Mark Sullivan

Environmental Devices Corporation
4 Wilder Drive Building #15
Plaistow, NH 03865
ISO-9001 Certified

Certificate of Calibration

Certificate Procedure Number: EDCQP200-4.11.5

Environmental Devices Corporation certifies the Haz-Dust Particulate Monitors are calibrated gravimetrically against the specifications and protocols set forth in NIOSH method 0600 and/or 0500. Calibration is NIST traceable and conforms to original published specifications of +/- 10%.

Calibration Dust Specifications are determined with a NIST traceable Coulter Mutisizer, ISO12103 –1 A2 Fine Test Dust and primary Flow Standard: LFE774300.

Our quality system standard meets the requirements of ANSI/NCSLZ540.1 and ASQC standard ISO 9001, MIL-STD 45662A, and customer's specifications if requested.

Calibration Test Dust Particulate	**Cumulative Volume Numeric Data**
Micron Size	% Less Than
1	2.9
2	11
3	19.6
4	27.7
5	34.6
7	43.6
10	52.1
20	70.7
40	89.2
80	99.8
120	100

Temperature = 22°C
Relative Humidity = 30%
Atmospheric Pressure = 760 mmHg
Measurement Uncertainty Estimated @ 95% Confidence Level (k=2)

Model	Serial Number	Calibration Date	Next Calibration Due Date
EPAM-5000	0714498	July 1 2014	July 2015

Calibration Span Accessory if purchased	K= 17.7 mg/m³	Model: CS-105

Technician	**Supervisor**
Matt Gosslin MG	Mark Sullivan

Environmental Devices Corporation
4 Wilder Drive Building #15
Plaistow, NH 03865
ISO-9001 Certified

PARTICULATES NOT OTHERWISE REGULATED, RESPIRABLE 0600

DEFINITION: aerosol collected by sampler with 4-µm median cut point CAS: None RTECS: None

METHOD: 0600, Issue 3 **EVALUATION: FULL** **Issue 1: 15 February 1984**
Issue 3: 15 January 1998

OSHA: 5 mg/m³
NIOSH: no REL
ACGIH: 3 mg/m³

PROPERTIES: contains no asbestos and quartz less than 1%; penetrates non-ciliated portions of respiratory system

SYNONYMS: nuisance dusts; particulates not otherwise classified

SAMPLING

SAMPLER: CYCLONE + FILTER (10-mm nylon cyclone, Higgins-Dewell [HD] cyclone, or Aluminum cyclone + tared 5-µm PVC membrane)

FLOW RATE: nylon cyclone: 1.7 L/min
HD cyclone: 2.2 L/min
Al cyclone: 2.5 L/min

VOL-MIN: 20 L @ 5 mg/m³
-MAX: 400 L

SHIPMENT: routine

SAMPLE STABILITY: stable

BLANKS: 2 to 10 field blanks per set

ACCURACY

RANGE STUDIED: 0.5 to 10 mg/m³ (lab and field)

BIAS: dependent on dust size distribution [1]

OVERALL PRECISION ($\hat{S}_{rT}$): dependent on size distribution [1,2]

ACCURACY: dependent on size distribution [1]

MEASUREMENT

TECHNIQUE: GRAVIMETRIC (FILTER WEIGHT)

ANALYTE: mass of respirable dust fraction

BALANCE: 0.001 mg sensitivity; use same balance before and after sample collection

CALIBRATION: National Institute of Standards and Technology Class S-1.1 or ASTM Class 1 weights

RANGE: 0.1 to 2 mg per sample

ESTIMATED LOD: 0.03 mg per sample

PRECISION: <10 µg with 0.001 mg sensitivity balance; <70 µg with 0.01 mg sensitivity balance [3]

APPLICABILITY: The working range is 0.5 to 10 mg/m³ for a 200-L air sample. The method measures the mass concentration of any non-volatile respirable dust. In addition to inert dusts [4], the method has been recommended for respirable coal dust. The method is biased in light of the recently adopted international definition of respirable dust, e.g., ≈ +7% bias for non-diesel, coal mine dust [5].

INTERFERENCES: Larger than respirable particles (over 10 µm) have been found in some cases by microscopic analysis of cyclone filters. Over-sized particles in samples are known to be caused by inverting the cyclone assembly. Heavy dust loadings, fibers, and water-saturated dusts also interfere with the cyclone's size-selective properties. The use of conductive samplers is recommended to minimize particle charge effects.

OTHER METHODS: This method is based on and replaces Sampling Data Sheet #29.02 [6].

TYPICAL - BACK SIDE OF CALIBRATION